

(12) INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(19) World Intellectual Property Organization
International Bureau



(43) International Publication Date
15 March 2001 (15.03.2001)

PCT

(10) International Publication Number
WO 01/17419 A1

(51) International Patent Classification⁷: A61B 5/00

(21) International Application Number: PCT/NO00/00289

(22) International Filing Date:
6 September 2000 (06.09.2000)

(25) Filing Language: Norwegian

(26) Publication Language: English

(30) Priority Data:
19994344 7 September 1999 (07.09.1999) NO

(71) Applicant (for all designated States except US):
LAERDAL MEDICAL AS [NO/NO]; P.O. Box 377,
N-4001 Stavanger (NO).

(72) Inventors; and

(75) Inventors/Applicants (for US only): MYKLEBUST,

Helge [NO/NO]; Solandsbakken 47, N-4025 Stavanger
(NO). EFTESTØL, Trygve [NO/NO]; Oddahagen 1,
N-4033 Forus (NO).


(74) Agent: PROTECTOR INTELLECTUAL PROPERTY
CONSULTANTS AS; P.O. Box 5074 Majorstua, N-0301
Oslo (NO).

(81) Designated States (*national*): AE, AG, AL, AM, AT, AU,
AZ, BA, BB, BG, BR, BY, BZ, CA, CH, CN, CR, CU, CZ,
DE, DK, DM, DZ, EE, ES, FI, GB, GD, GE, GH, GM, HR,
HU, ID, IL, IN, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR,
LS, LT, LU, LV, MA, MD, MG, MK, MN, MW, MX, MZ,
NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, SL, TJ, TM,
TR, TT, TZ, UA, UG, US, UZ, VN, YU, ZA, ZW.

(84) Designated States (*regional*): ARIPO patent (GH, GM,
KE, LS, MW, MZ, SD, SL, SZ, TZ, UG, ZW), Eurasian
patent (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European
patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE,

[Continued on next page]

(54) Title: SYSTEM FOR PREDICTING THE OUTCOME OF AN IMAGINARY DEFIBRILLATOR SHOCK



(57) Abstract: A system for calculating a probability figure for the outcome of an immediate following defibrillator shock, containing a device (3) for measuring and digitising bio-electrical signals that characterise the heart, is connected to an analysis unit (2) that organises the signals into segments on a continuous basis. The analysis unit (2) includes means of calculating the property vector for each of the segments. The analysis unit (2) quantifies the probability figure of the property vector of at least the one class predicting at least one of several possible outcome classes of an immediate following defibrillator shock. Furthermore, the analysis unit is connected to a display unit that presents the probability figure, and/or the analysis unit is connected to a decision support system that utilises the probability figure when selecting the type of treatment.

WO 01/17419 A1



IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GW, ML, MR, NE, SN, TD, TG).

For two-letter codes and other abbreviations, refer to the "Guidance Notes on Codes and Abbreviations" appearing at the beginning of each regular issue of the PCT Gazette.

Published:

- *With international search report.*
- *Before the expiration of the time limit for amending the claims and to be republished in the event of receipt of amendments.*

SYSTEM FOR PREDICTING THE OUTCOME OF AN IMAGINARY DEFIBRILLATOR SHOCK

The present invention regards a system for calculating the probability figure for the
5 outcome of an immediate following defibrillator shock on the basis of properties of the
heart measured during sudden cardiac arrest and resuscitation, as stated in the
introduction to Claim 1.

Near 40% of all those who suffer sudden cardiac arrest will be able to survive if they
10 receive good, lifesaving treatment immediately. When treatment is delayed, the chances
of survival decrease, cf. the article by Holmberg S, Holmberg M: "National register of
sudden cardiac arrest outside of hospitals " 1998 [1]. The treatment primarily consists of
cardio-pulmonary resuscitation (CPR), which is administered until a defibrillator is in
place. Thereafter, the treatment consists of alternating use of a defibrillator and CPR
15 until resuscitation or until an ALS team arrives. (ALS = "Advanced Life Support") The
latter also includes medication and securing of the respiratory passages as part of the
treatment, cf. ILCOR, "Advisory statements of the International Liaison Committee on
Resuscitation. Circulation" 1997;95:2172-2184 [6]

20 Scientific papers in recent years point out a number of factors that affect the chances of
survival:

Time: The chance of surviving sudden cardiac arrest falls with time from
heart failure until the first defibrillator shock is administered.[1]

CPR: The chance of surviving increases when someone administers CPR
before the defibrillator arrives.[1]

Quality of Studies show that the quality of the CPR influences the survival.

CPR: (Cf. the publications by Wik L, Steen PA, Bircher NG. "Quality of
bystander CPR influences outcome after prehospital cardiac arrest.
Resuscitation" 1994;28:195-203[2].

Gallagher EJ, Lombardi G, Gennis P. "Effectiveness of bystander
CPR and survival following out-of-hospital cardiac arrest". J Am
Med Assoc 1995;274:1922-5[3]

Van Hoyvegen RJ, Bossaert H. "Quality and efficiency of

bystander CPR. Resuscitation" 1993;26:47-52[4])

Timing of A study shows that when the duration of sudden cardiac arrest
CPR and exceeds a number of minutes, the chance of survival will increase if
defibrillator the ambulance personnel first administers a period of CPR before the
treatment: defibrillator is used. (Cf. Cobb L, et al. "Influence of cardiopulmonary
resuscitation in patients with out-of hospital ventricular fibrillation.
JAMA", April 7, 1999 – Vol 281, No 13 [5])

In case of sudden cardiac arrest, the electrical activity in the heart (ECG) will indicate
the state of the heart. Today's defibrillators measure and analyse ECG in order to
classify the rhythm. If the rhythm is classified as Ventricle Tachyardia (VT) or
5 Ventricle Fibrillation (VF), defibrillator treatment may have an effect. VT is often the
precursor of VF. VF will as time goes by and the energy and oxygen reserves of the
heart muscle are depleted tend towards Asystole, a rhythm characterised by very little or
no electrical activity. The purpose of the defibrillator treatment is to restore the
organised electrical activity of the heart and the associated blood pressure and blood
10 circulation. This is often denoted ROSC – "Return of Spontaneous Circulation"; and is
the first step towards survival.

Only a fraction of the shocks delivered actually result in ROSC. Most shocks today do
not give ROSC, cf. the publications Gliner BE et al. "Treatment of out-of hospital
15 cardiac arrest with a Low-Energy Impedance-Compensating Biphasic Waveform
Automated External Defibrillator" [7], Sunde K, Eftestøl T, Askenberg C, Steen PA.
"Quality evaluation of defibrillation and ALS using the registration module from the
defibrillator. Resuscitation" 1999 [14]. In general, it can be said that the chance of
ROSC is at its greatest immediately after sudden cardiac arrest, when the heart muscle
20 still possesses energy reserves and oxygen. Many patients achieve ROSC after
alternating use of shocks and CPR. The disadvantages of having to give many shocks
are several: First of all, no CPR will be given during the shock treatment, a factor that
further aggravates the situation for the vital organs, particularly for the brain.
Furthermore, it has been shown that the heart muscle is also damaged by the shocks,

and that the damage increases with the number of shocks and the amount of energy, cf. the publications Ewy GA, Taren D Bangert J et al. "Comparison of myocardial damage from defibrillator discharges at various dosages. Medical instrumentation" 1980;14:9-12. [16]. For the patient, the ideal would be to be given only one shock, and for this
5 shock to give ROSC.

Thus, for many patients, it is crucial that the administration of CPR be effective, so as to revitalise the heart through supplying a flow of blood through the heart muscle, cf. the publication Michael JR et al. "Mechanism by which augments cerebral and myocardial
10 perfusion during cardiopulmonary resuscitation in dogs. Circulation" 1984;69:822-835. [17]. This revitalisation can be indicated through ECG measurements, where ECG characteristics such as form, spectral flatness measurements, frequency, amplitude, energy etc. is seen to change back towards their initial values such as they would have been immediately after the heart action was suspended, cf. the publications Eftestøl T,
15 Aase, SO, Husøy JH. "Spectral flatness measure for characterising changes in cardiac arrhythmias." Computers in Cardiology, [15] and Noc M, Weil MH, Gazmuri SS, Biscera J and Tang W. "Ventricular fibrillation voltage as a monitor of the effectiveness of cardiopulmonary resuscitation" in J Lab Clin Med, September 1994 [13]. This revitalisation will increase the probability of the next shock giving ROSC.

20

Unfortunately, not everyone survives. For many, the reason behind the sudden cardiac arrest is such that resuscitation is impossible. Furthermore, the time factor and the quality of the treatment will also play a part and affect the survival.

25 Today's resuscitation guidelines describe a protocol that is the same for everyone, regardless of sex, race, how long the heart action has been suspended, whether a member of the public has given CPR etc. The means of resuscitation are primarily CPR and defibrillator treatment, and later also medication administered by lifesavers who have been given special training in this area. Today's protocol is such that if the first
30 three shocks have no effect, CPR is to be given for 1 minute, then three more shocks, and so on. As it takes about one minute to give three shocks, the patient will be without CPR for half of the time.

Literature and other patent applications describe technology, the object of which is to guide the lifesaver in the choice between CPR and defibrillator treatment. Brown et al in US patents no. 5 683 424 and 5 571 142 [10] describe a system that, based on spectral
5 measures in VF, instructs the lifesaver to either give CPR or give a shock. A separate analysis of this method, where the method has been tested on human VF, yields results that show the method to have a low specificity, i.e. that the method will only to a limited degree reduce the number of unnecessary shocks. Noc M, Weil MH, Tang W, Sun S, Pernat A, Bisera J. "Electrocardiographic prediction of the success of cardiac
10 resuscitation" in Crit Care Med 1999 Vol 27 No 4 [12] describe a similar system, based on an animal model, which links the mean amplitude and dominant frequency of VF to the outcome of the defibrillator shock. Both of these methods aim to advise against defibrillator use as long as the condition of the heart is such that a shock is assumed not to have an effect, and instead use CPR. Both methods define absolute criteria based on a
15 limited number of observations from a defined group of patients or animals.

The object of the present invention is to seek to constantly optimise the treatment through:

- 20 a) By means of equipment connected to the patient, measuring the properties of the heart during sudden cardiac arrest, and, based on relevant treatment, knowledge of the patient and knowledge of comparable conditions and outcomes of treatment carried out previously, calculate the probability figure of achieving ROSC – Prosc.
- b) Presenting the probability figure or using the probability figure in support of the
25 decision on further treatment.
- c) Relaying registrations for each treatment, along with the outcome of each shock, to a centrally located computer, and using this experience to improve the basis for Prosc calculations, so as to improve the confidence of the next calculation.
- d) Possibly registering the measured CPR parameters and, together with
30 information regarding medication given, looking at the development of the probability figure for the purpose of identifying effective lifesaving.

- e) Possibly providing feedback to the user regarding which CPR parameters and what medications have been identified as effective, or alternatively through using this information in order to instruct the user in effective CPR.
- 5 To the extent the word "possibly" is used, this is done in order to take into consideration the fact that there are many categories of users and lifesavers, and based on their professional level and economy, it is appropriate to use the solutions that best serve the purpose.
- 10 The object of the invention is to contribute towards giving the patient a treatment that is better suited to the individual, and which gives a greater chance of survival. The use of empirical data makes it possible to allow for differences and the ever-changing patient groups and treatment. To the extent that not all shocks give ROSC, the system attempts to take into consideration factors that affect or may affect the probability of ROSC:
- 15 • Properties of the heart that may be observed via the electrodes, and which express the metabolic condition and the pumping action. These properties develop with time during sudden cardiac arrest, but may be partially reversed through the use of medication and CPR.
 - Type of defibrillator shock and energy selection produce varying effectiveness.
 - 20 • Physical conditions. A big patient will receive a lower current density through the heart than a small patient, for the same energy selection.
 - Patient information. The system takes into account the fact that there may be a difference between men and women, based on the fact that over 70% of those who suffer sudden cardiac arrest are men. In certain parts of the world, the
 - 25 duration of life has increased, so that the number of elderly who suffer sudden cardiac arrest is increasing. These may very well have received treatment over a period of time, both medications, surgical procedures and aids such as pacemakers, all of which may have an effect on the Prosc.
 - Geography, race. The system further takes into consideration the fact that there
 - 30 may be differences based on lifestyle and genetic conditions, just as life expectancy varies greatly depending on geography and race.

The above is provided by means of a system of the type mentioned initially, the characteristics of which are stated in Claim 1. Further characteristics of the invention appear from the remaining, dependent claims

- 5 Using Prosc to optimise the treatment may be done in several ways. For advanced users, the most expedient will be to present the parameter graphically versus time, as a trend curve. This will immediately provide a direct indication of the state of the heart, and also indicate the effect of medication and CPR. For groups who are not trained in relating to this type of information, the most appropriate thing will be to provide
- 10 automatic decision support in the question of whether or not to give CPR, in which way CPR should be given, or whether shocks should be given. The principle of simple decision support may be:
- If Prosc is less than a limit, Grosc, CPR is recommended. Otherwise, a number of sequential defibrillator shocks are recommended.
 - 15 • CPR is recommended until positive changes in the Prosc level out, but no further than T minutes before a number of defibrillator shocks are recommended.

The following will describe the invention in greater detail, with reference to the drawings, in which:

- 20 Fig. 1 shows system components consisting of one, alternatively several, computers in a network that communicates with a number of positioned analysis units.
- Fig. 2 shows the block diagram for a defibrillator with a built-in analysis unit.
- Fig. 3 shows the elementary flow diagram for information.
- 25 Fig. 4 shows an apparatus with electrodes connected to the patient's chest, in positions on the chest that are normally used for delivery of a defibrillator shock, as well as for measuring ECG in accordance with standard derivation II.
- Fig. 5 shows a flow diagram for development of coefficients to optimal filter.
- 30 Fig. 6 shows a flow diagram for development of a classification that fulfil the requirement of generality.
- Fig. 7 shows a general block diagram of the invention with focus on the analysis unit.

The system consists of one, alternatively several, computer(s) 1 in a network that can communicate with a number of positioned analysis units 2. These may either be integrated into equipment (U1, U2..) such as defibrillators or ECG monitors, or they may occur in or as a support product used during the resuscitation attempt. The analysis
5 units 2 generally operate independently of the computers 1, however after use, the analysis units will deliver field data to the computer 1, and possibly receive adjusted algorithms for calculation of property vector and/or Prosc

The analysis unit 2 normally has subsystems associated with it, cf. Fig. 2:

10 Some of these subsystems are standard in equipment such as defibrillators and ECG monitors, and these are as follows:

Reference number 3 denotes a system for measuring ECG, connected to electrodes E. For measuring and digitalize of bio-electrical signals is used said electrodes connected to skin of the patient in accordance with standard derivation II, cf. Fig. 4, which shows
15 the device with electrodes connected to the patient's chest in locations that are normally used for delivering defibrillator shocks, as well as for measuring ECG in accordance with standard derivation II. In this derivation, ECG will essentially express the electrical activity in the longitudinal axis of the heart. Reference number 4 denotes a system for checking the electrode connection to the patient, reference number 5 denotes a device
20 for high voltage and shock delivery, if integrated into a defibrillator, reference number 7 denotes an algorithm for classification of ECG, if integrated into a defibrillator, reference number 7 denotes a microprocessor system and software, reference number 8 denotes computer memory, reference number 9 denotes a user interface, reference number 10 denotes an energy supply, reference number 11 denotes a system for delivery
25 and receipt of data from external equipment.

Subsystems 3-11 are standard equipment in defibrillators and ECG monitors, and will therefore not be described further in this specification.

30 Analysis unit 2 consists of the following units:

Unit 12 for determining one or more properties of the heart that is combined in a property vector, and based on this calculate the probability of ROSC, Prosc, for the patient who is connected up. Possibly also a module 13 for determining the blood flow through the heart, based on the measured impedance and the change of the impedance
5 between the electrodes as a function of the pumping action of the heart and the expansion of the lungs, a possible module 14 for registering CPR parameters from sensors S, a possible module 15 for registering patient specific information, a possible module 16 for registering any medication administered; and a possible module 17 for correlating positive changes in Prosc with information regarding the treatment given,
10 and display or use this information to guide the treatment.

The analysis unit may be realised as a self-contained microprocessor unit, or it may be realised through the microprocessor unit in the equipment.

15 Each computer consists of the following subsystems:

(a) Hardware, (b) operating system, (c) software and interface for communication in a network (d) database for field data, (e) algorithm for calculation of property vector, including a classification system for predict the outcome of an immediate following
20 defibrillator shock, (f) system for calculation of Prosc, (g) algorithm for correlating changes in Prosc with information regarding patient and treatment, and (h) system for delivery and receipt of data from positioned defibrillators.

With regard to the computers, the subsystems of hardware, operating systems, software
25 and interface are of a generic nature, and will not be described in greater detail.

A more detailed description of the system, first computer 1.

(d) The database for field data consists of a large amount of patient episode data, and
30 contains:

- Patient information: Sex, age, weight, race etc.
- Geographical information

- Information regarding each defibrillator shock: Curve shape, energy, timing versus VF.
- For each shock:
 - Preshock ECG
 - 5 - Preshock CPR data
 - Preshock medication data
 - Preshock impedance data
 - Postshock ECG
 - Postshock impedance data
 - 10 - Annotation of ROSC/Non-ROSC, with outcome rhythm for each shock

(e) The algorithm for calculation of the property vector (v) makes use of mathematical methods in order to characterise the condition of the heart based on a recording of a bio-medical signal (x). The bio-medical signal is preferably ECG, but also other type of
 15 signals such as signals derived from the change of the impedance, sound or pattern of the movement may be considered.

The algorithm for calculation of the property vector is hereafter denoted as $v(x)$. $v(x)$ which is used on ECG from empirical data provide us two set of property vectors:
 20 An set, $V1$, containing $n1$ property vectors where the outcome of the shock is ROSC, and an set, $V2$, containing $n2$ property vectors where the outcome of the shock is none-ROSC.

In general, $v(x)$ is defined as an operator that operates on an ECG sequence, x ,
 25 consisting of N samples, which generates a property vector, v , consisting of M vector elements that ideally takes care of the information in x that separates the group of x that results in ROSC, $X1$, from the group of x that results in non-ROSC, $X2$. The methods of property extraction are innumerable, and the literature describes some of these, which can be roughly divided into time and transform domain methods, where the object is to
 30 structure x in a manner that is appropriate for property extraction. Among preferred methods are:

- Optimised digital filters determined by L filter parameters that divides x into M channels. The energy from each of these channels is calculated, so as to make the property vector consist of M elements. These types of filters are described inter alia by T. Randen "Filter and Filter Bank Design for Image Texture Recognition" in a thesis of NTNU, Oktober 1997 where the filters are optimised in order to achieve the best possible recognition of the different textures. For present purpose is the optimised filters found by using a numerical gradient search algorithm (T. Coleman, M. A. Branch and A. Grace, , *Optimization Toolbox for Use with MATLAB*, The Math Works Inc, 1999) to achieve the best possible separation of the ROSC group from the non-ROSC group. Separation ability is measured by the sum of sensitivity (degree of correct recognition of ROSC) and specificity (degree of correct recognition of non-ROSC) By a given iteration in optimisation is this performance measured, and the set of parameters, which define the filters, is adjusted in the direction corresponding the increase in performance. This procedure is repeated until maximum performance is reached.
- Spectral measures that are calculated on the basis of the estimate of the power density spectrum (PSD) of x . The PSD can be estimated through use of Fourier transforms. Based on the PSD, characteristics are calculated that express the frequencies at the centre and the maximum points of the PSD. In addition, the flatness and energy of the PSD is also characterised

Examples of other methods for property extraction are:

- wavelet analysis,
- neural networks

25

The relation between $V1$ and $X1$, $V2$ and $X2$ respectively are as following: $X1$ containing an set of $n1$ ECG sequences, which, when used on $v(x)$, provides a number of property vectors $V1$, which all is belonging to the outcome class ROSC ($w1$). $X2$ containing an set of $n2$ ECG sequences, which, when used on $v(x)$, provides a number of property vectors $V1$, which all is belonging to the outcome class non-ROSC ($w2$).

30

(f) A system for calculation of the Prosc function is based on pattern recognition theory, and forms the second element of the classification system. In this context, the term classes is defined as the collection of measurements of the condition of the heart that corresponds to

- 5 - ROSC ($w1$)
- non-ROSC ($w2$)

The property vectors of the two classes are statistically described by

- 10 - $P(wi)$, $i=1,2$, which is the a priori probability of the two classes. I.e. before a measurement is made, the probability of one or the other outcome is known through the respective a priori probabilities.
- $p(v|wi)$ are the class specific probability density functions. These express how the measurements within the given classes are distributed. $p(v)$ expresses the compound probability density function for the measurements, and is given by adding up the class specific probability density functions weighted by the associated a posteriori probabilities.
- 15 - $P(wi|v)$ are the a posteriori probability functions for the two classes. These functions express the probability of a given measurement belonging to wi . Bayes formula expresses $P(wi|v)$ as a function of the above probability functions .
- 20 - $P(wi|v) = P(wi) * p(v|wi) / (p(w1) + P(w2) * p(v|w2))$
- The sum of the a posteriori probabilities for a given v is always 1.
- 25 In the case of a given measurement, v , one wishes to determine the class allocation $w1$ or $w2$. It has been proven that the expected probability of misclassification is minimised by selecting the wi that corresponds to the maximum $P(wi|v)$. It is further possible to define (make an estimated choice of) the cost of all types of misclassification, such that the expected risk of a given misclassification is given by the product of the cost and the
- 30 a posteriori probability of the true class. The expected risk of misclassification can then be minimised by classification is a class corresponding the product with smallest value is selected.

In most cases, the statistics of the property vector are not known. These quantities must then be estimated before $\text{Prosc}(v)$ can be produced. The pattern recognition theory describes a multitude of methods for this, which are based on measurements (practice data) that are examples from the various w_i . Some examples:

- 5 - Histogram techniques, which divides the outcome space into hypercubes in which the probabilities within each of these are calculated on the basis of the number of occurrences of the different classes within the given hypercube. This corresponds to the method used herein. In the following is described how statistic quantity is estimated.

10 We will start defining the quantities:

n = total number of observations in the empirical material.

n_1 = total number of observations corresponding ROSC outcome.

n_2 = total number of observations corresponding non-ROSC outcome.

15 n_{j1} = total number of observations corresponding ROSC outcome within hypercube no. j .

n_{j2} = total number of observations corresponding non-ROSC outcome within hypercube no. j

We have $n = n_1 + n_2$. Estimate for a priori probability will then be

20 $\hat{P}(w_i) = n_i/n, i=1,2.$

The local estimates (within hypercube j) for the class specific probability function will then be

$\hat{p}(v|w_i) = n_{ji}/n_i, i=1,2.$

25 The local estimates for a posteriori probabilities is calculated in respect of the Bayes formula inserted estimate for a priori probability and the local class specific probability density functions. Se R. J. Schalkoff.

Pattern recognition: Statistical, structural and neural approaches. John Wiley & sons, New York (NY), 1992

$\hat{P}(w_i|v) = n_{ji}/(n_{j1} + n_{j2}), i=1,2$

- 30 - Radial base functions, in which the probabilities at a given point are calculated on the basis of the contribution from surrounding practice data from the different classes. The contributions decrease with distance.

- Parametric modelling, in which a mean value and dispersion for the different classes are used to produce analytical probability models.
- Neural networks, learning vector quantization and nearest neighbour classification are some other central methods within the pattern recognition theory.

It is important that a given classifier be tested on a set of observations (test set) independently of the practice data (practice set), in order to check that the classifier yields the expected results. The demand is that there is consistency between practice and testing, that the classifier fulfils the requirement of generality. With generality means: By dividing the empirical data in two parts and letting the one part represent a set of data called practice set and the other part represent a set of data called test set, the generality is defined as following: Decision limits used on all of the property vectors in each set of data for classification of the outcome, which provides approximately same performance (the sum of sensitivity and specificity) for both set of data fulfil the requirement of generality. These decision limits occur through an iterative process where the practice set is included in the calculation of the decision limit, see fig. 6.

Those measurements v that correspond to the ROSC outcome belong in wI . The probability of a given measurement, v , belonging in wI is given by $P(wI|v)$. In other words, this probability function expresses the probability figure $Prosc$ of ROSC for a given measurement v .

$$Prosc(v) = P(wI|v)$$

As mentioned previously, different property vectors, v , can be calculated by means of a countless number of methods. Which methods and which dimension, M , is suitable for expressing $Prosc(v)$ is assessed on the basis of the expected risk in the case of misclassification for each method. The method that minimises this risk is the most appropriate for expressing $Prosc(v)$. Fig.6 shows a flow diagram for an iterative development of algorithm for calculation of the property vector v . Basis for the iterative development is empirical data. As the amount of empirical data increase is this iterative

process repeated so that the ability of the property vector to predict outcome classes is increased. The iterative adjustment of the decision limit is also included so that the requirement of generality is fulfilled.

- 5 (g) The algorithm for correlating changes in Prosc with information regarding the patient and the treatment is mainly for scientific purposes. The defibrillator to guide the user during lifesaving may later use the results from the correlation.

Prosc(v) has been provided as described under points (d) and (e). In this analysis, ECG
10 sequences are extracted from the patient material, so that the ECG sequences describe a course of treatment that is as uniform as possible, with the complete patient material seen under one. Examples of such a course of treatment may be

- CPR sequences
- "Hands off" intervals, for instance a period for defibrillator rhythm
15 analysis up to the shock, after a CPR period.

In these ECG sequences, corresponding Prosc(v) sequences are calculated as described under points (d) and (e). Consequently, the change in Prosc(v), DProsc, is calculated for each sequence. DProsc is grouped on the basis of those treatment characteristics that
20 are of interest with regard to the effect of the treatment. As an example, one can group DProsc with regard to the following treatment characteristics, singly or in combination:

- Different compression frequencies, compression depths, duration of chest compression
- Degree of ventilation
- 25 - Medication
- Physiological measurements such as blood flow measurements, blood pressure etc.

Where significant differences in DProsc occur for dissimilar treatment conditions, this information may be used to identify advantageous treatment methods. This information
30 may be utilised through the person giving the treatment being given feedback regarding good and poor treatment.

(h) A system for delivering and receiving data from positioned analysis units. Here, no special requirements apply. The exchange of data can take place directly through use of memory modules such as PCMCIA, cordlessly by means of IR or RF communication, via networks such as the Internet, or by a direct connection between communication
 5 ports in the equipment and the computer. The most practical method these days is to have the analysis unit 2 communicate directly with the computer 1 via a local PC that it can communicate with, and to have the local computer pass the data on via the Internet.

Detailed description of the analysis unit 2.

10

System 12, algorithm $v(x)$ for calculating the property vector (v) and algorithm for calculating the probability of ROSC, $Prosc$, for ECG from the patient who is connected up.

- 15 - $v(x)$ is a set with calculation, which combined form a property vector. The calculation is a set of energy calculation within a determined frequency band (optimised filter) or a set of parameters diverted from effect density spectrum or a combination of this.
- 20 - The algorithm for calculation of $Prosc$ is typical a matrix where the number of dimension is consistent with the number of dimensions within the property vector, where each matrix element is containing a numerical value for $pROSC$. The numerical value is downloaded from the computer 1. The matrix is provided in that there for each ECG sequens within the empirical material is calculated a property vector. Property vectors that are approximately identical are grouped thereafter together in a number
 25 of sets, where each set is assigned a matrix element. For an set, j , the occurrences of ROSC is counted, n_{j1} , towards the occurrences of non-ROSC, n_{j2} . The ratio between the numbers of {ROSC} towards the number of {ROSC+non-ROSC} define the numerical value of $pROSC$ for the set j which is $n_{j1}/(n_{j1}+n_{j2})$. This correspond the local estimates
 30 for a posteriori probability for the class w_1 as described by use of histogram technique. The numerical value for each set is then stored in

that matrix element that correspond to the elements of the property vector.

The module 13 for calculating blood flow through the heart based on the measured impedance and the change of the impedance between the electrodes as a function of the pumping action of the heart and the expansion of the lungs:

- The value of the measured impedance, Z_0 , measured by means of an approximately constant alternating current, informs the analysis unit 1 of the impedance between the electrodes, and can be used to replace system 4.
- The impedance change between the electrodes will be proportional to the change in the set of air in the lungs plus the working volume of the heart. The change due to air dominates. By looking at the signal between two ventilations, or by first filtering out the ventilation, it will be possible to estimate the working volume on the basis of the formula

$$\Delta V = \frac{\Delta Z \rho L^2}{Z_0^2}$$

20

This formula is universally known, and is used within Impedance Cardiography.

ΔZ is the impedance change, ρ is the resistivity of the blood, L is the distance between the electrodes, and Z_0 is the numerical value of the impedance. A simplification of this formula is preferable:

25

$$\Delta V = \frac{\Delta Z k}{Z_0^2}$$

Here, k is a constant. This measurement will indicate to what degree the blood is flowing, and will contribute towards characterising the condition of the heart in VF/VT, and will furthermore indicate ROSC in case of a successful defibrillator shock.

Module 14 for measuring and registering CPR parameters.

Relevant CPR parameters are:

- Inflation time and inflation volume are measured by looking at the impedance change between the electrodes. This change is several times greater than the change that takes place as a function of the blood stream from the heart, and is proportional to the amount of air in the lungs. The principle is known from other diagnostic equipment.
- Compression rate and compression proportion (the ratio between compression time and relaxation time) during chest compression may be measured by looking at the impedance change between the electrodes, or by placing a sensor at the compression point on the patient's chest. This sensor can contain a pressure switch, or a dynamometer or an accelerometer.
- Compression depth is calculated on the basis of signals from an accelerometer placed at the compression point.
- Time between inflation and chest compression, and time between chest compression and inflation.
- Proportion of CPR relative to the total treatment time
- Amount of compression, the sum of the product between the duration and depth of the compression.

Module 15 for indicating patient specific information.

This information can be passed to the analysis unit 1 e.g. by dedicated push buttons, or it may come in from an external source such as a patient database or a patient journal on a PC / handheld computer. Relevant information is:

- Geographical area
- Age
- Sex
- Weight
- Race

Module 16 for indicating medication and dosage given.

This information can be passed on to the analysis unit 1 through dedicated push buttons, from a patient journal on a PC or from other devices that log the use of medication.

Relevant medicines are

- Epinephrine
- 5 - Lidocaine
- Bretylium
- Magnesium sulphate
- Procainamide
- Vasopressins
- 10 - Thrombolysis medication
- and so on.

Module 17 for correlating positive changes in Prosc with information regarding the treatment given, and displaying or using this information to guide the treatment.

- 15 - The system identifies and registers a period of Prosc with positive change. At the same time, the system identifies and registers the average of each CPR parameter measured for a period of time prior to the change and during the change, and if applicable, what medication was given during the same period.
- 20 - If a new period of improved positive change occurs, the identification of CPR parameters and medication is repeated.
- This information can be displayed on the defibrillator screen, or it may be used to produce voice messages that guide the user to deliver CPR with parameters that are in accordance with the registration.
- 25 - This information will also be of great importance to research, with a view to optimising the guidelines for CPR treatment and training.

To summarise, the invention can be described with reference to Figure 3, which schematically shows the flow of information between the central computer 1 and the
30 analysis unit 2 in the positioned equipment in an application of the invention.

The computer 1 contains empirical data from previous resuscitation attempts, where the outcome of the resuscitation attempt is known. The main ingredient in the empirical

basis is the ECG and the associated outcome after a shock (ROSC/non-ROSC). Additional empirical data impart nuances to the relationship between outcome, treatment and patient specific factors.

- 5 An algorithm $v(x)$ for calculating of a property vector v is developed based on empirical material. This algorithm is adapted so that the ability of the property vector to predict the outcome (ROSC/non-ROSC) is optimised within the demand of generality.

A property vector v is thus characterised by that it represents ECG segments that is
 10 succeeded by ROSC after shock characteristically different from ECG segments that is succeeded by non-ROSC after shock. Such property vector is provided through an iterative process where the parameters for each iterative step is being adjusted until the best performances are reached, see fig. 5.

- 15 By grouping the property vectors in the empirical data set that are approximately identical and adding up the number of cases of ROSC and the number of cases of non-ROSC for each of these sets, then the probability figure $Prosc$ for ROSC is expressed statistically relative to the value of an in-coming property vector which represents the ECG segment $Prosc$ is desired to be calculated for.

20 The property vector of a ECG segment consist of a number of element which together characterize the ECG segment with a view to the outcome after a shock. Let the property vector consist of $(n+p)$ elements that together describes one determined ECG segment. Then the property vector will be n energy calculations of the output signal of n
 25 different digital filters, where the ECG segment constitute the input signal plus p parameters that express characterizing features by the power density spectrum of the ECG segment. Preferred characterizing features is: Frequency by the centre of gravity, frequency by the maximum point, spectral flatness measurements, and spectral energy. These parameters will to a limited degree overlap each other in view of information
 30 contents. In order to reduce the set of data without to reduce the number of information may for example be used principal component analysis. By such a method may a big set of data, which is described of $(n+p)$ elements in each vector be reduced to one set of data where each vector only have m elements; $m < (n+p)$.

- 35 A practical way of expressing this statistical interrelationship is through a $Prosc$ function, which is a substitute for all the empirical data, but which mathematically expresses the same relationship between the property vector and $Prosc$.

This function is entered into the programme code of the analysis unit, so that when this receives a segment of ECG, the analysis unit will first perform the same calculation of the property vector as that performed by the computer, and then use the property vector
5 as input to the Prosc function in order to calculate the probability figure of an immediate following defibrillator shock giving ROSC.

The analysis unit may furthermore impart nuances to, and thereby provide a more accurate probability figure, by also utilising knowledge of the treatment and the patient,
10 seen in relation to empirical data.

The ever-changing forms of treatment and patient characteristics necessitate a continuous update of the empirical basis. This is achieved by each analysis unit passing on its experience regarding each shock to the central computer, where the central
15 computer repeats the grouping of the property vectors, recalculates the Prosc function and passes the result back to the analysis unit.

C l a i m s

1.

A system for calculating probability figures for the outcome of an immediate following
5 defibrillator shock resulting in return of spontaneous circulation (ROSC)

w h e r e i n

an analysis unit is connected with a module, which is measuring bio-electrical signals
from electrodes connected to a patient,

the analysis unit is provided to organise the bio-electrical signals continuous into
10 segments,

the analysis unit is provided for each segments to calculate a combination parameter
that characterise the condition of the heart

the analysis unit is provided to by means of comparing for each combination of
parameters to find a corresponding combination of parameters from earlier made
15 defibrillator treatments, where there for each combination of parameters is assigned a
probability figure, where the probability figure expresses the number of defibrillator
shocks that results in ROSC relative the total number of defibrillator shock for each
combination of parameters, and

the analysis unit has an output for the probability figure.

20

2.

A system according to Claim 1, w h e r e i n
the bio-electrical signals is ECG signals.

25 3.

A system according to Claim 1-2 w h e r e i n
the analysis unit (2) is provided to calculate the probability figure by means of an
algorithm.

30 4.

A system according to Claims 1-3, w h e r e i n the algorithm for
calculation the probability figure is a table look up in a m-dimensional table, where

there for each table element is stored a numerical value for the probability figure, the table look up is determined from the value of a m-dimensional vector, the value of the m-dimensional vector is diverted from the calculation of the energy of respective m different signal sequences that is represented on the output of m different digital filters, 5 where the signal on the input of each digital filter is the segment of the ECG signal.

5.

A system according to Claims 4, w h e r e i n the value of the m-dimensional vector is diverted from calculation of flatness, energy, 10 frequency by the centre of gravity and frequency by the maximum point of a power density spectrum, where the power density spectrum is diverted from the ECG signal segment.

6.

15 A system according to Claims 1-5, w h e r e i n the calculation unit is connected to a data storage, the calculation unit is storing for each treatment parameters which describe the patient and parameters which describes the treatment, the calculation unit is connected to means for exchange of data, the exchange of data occur on a regular basis towards a central computer, where the calculation unit receives optimised 20 algorithm for calculation of the probability figure, and the computer receives information that is stored en the data storage.

7.

A system according to Claims 1-6, w h e r e i n 25 there is provided an optimised algorithm by first establishing of an updated set of empirical data consisting of information from a number of new patient treatments together with information from a number of earlier performed patient treatments, which all contain sequences of ECG where the outcome after shock are known; the optimised algorithm occur by iterative search after filter coefficients by m digital filters, where the 30 filter coefficients are adjusted iterative in view of performance of a classification routine, where again the classification routine is adjusted iterative in view of performance and generality, where the performance is defined as the sum of sensitivity

and specificity for classification of each of the ECG sequences to outcome classes ROSC and non-ROSC respectively, where the real outcome of shock is known for each ECG sequence, generality is fulfilled as the classification routine has the same performance for a arbitrary composite half of empirical material as for the rest of the empirical material, where measurement of generality and performance is provided in that each ECG sequence in the empirical material is expressed as a m-dimensional vector calculated from energy at the output of m digital filters, where the classification routine classifies each m-dimensional vector to one of the outcome classes ROSC, non-ROSC respectively, where the performance is measured as the sum of sensitivity and specificity of the classification routine, where a arbitrary composite half of empirical material has the same performance as the rest of the empirical material, the optimised algorithm for calculation of the probability figure consist of a matrix having m matrix elements, where each matrix element express a probability figure, where the probability figure for each matrix element is provided by grouping ECG-sequences which is expressed by approximately identical m-dimensional vectors, where the occurrence of ECG which resulted in ROSC by shock plus the sum of occurrence of ECG which resulted in non-ROSC by shock constitute the probability figure for the matrix element, the m-dimensional matrix together with the filter coefficient constitute the optimised algorithm for calculation of the probability figure.

20

8

A system according to Claims 1 w h e r e i n
the output of the analysis unit is connected to a receiver in the shape of a display unit.

25 9.

A system according to Claim 8, w h e r e i n
the receiver is a defibrillator.

10.

30 A system according to Claims 1, w h e r e i n the receiver of the probability figure is an algorithm for decision support for the choice of treatment.

11.

A system according to Claims 1, w h e r e i n
the analysis unit (2) identifies periods of positive change in the probability figure
together with parameters that characterise the treatment, and passes on the numerical
5 value of the positive change in the probability figure, together with the mean value of
each treatment parameter over the period, to a receiver.

12.

A system according to Claims 1, w h e r e i n
10 the receiver of the numerical value of the positive change in the probability figure,
together with the mean value of each treatment parameter over the period, is a display
unit.

13.

15 A system according to Claims 1, w h e r e i n
the receiver of the numerical value of the positive change in the probability figure,
together with the mean value of each treatment parameter over the period, is an
algorithm for decision support for the choice of treatment.

20 14.

A system according to Claims 1-13, w h e r e i n
a device for indicating patient specific information and/or specific information
regarding the treatment is connected to the analysis unit (2).




Fig. 1




Fig. 2




Fig. 3




Fig. 4




Fig. 5




Fig. 6




Fig. 7

1
INTERNATIONAL SEARCH REPORT

International application No.
PCT/NO 00/00289

A. CLASSIFICATION OF SUBJECT MATTER IPC7: A61B 5/00 According to International Patent Classification (IPC) or to both national classification and IPC		
B. FIELDS SEARCHED Minimum documentation searched (classification system followed by classification symbols) IPC7: A61B Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched SE,DK,FI,NO classes as above Electronic data base consulted during the international search (name of data base and, where practicable, search terms used)		
C. DOCUMENTS CONSIDERED TO BE RELEVANT		
Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
P,X	US 5957856 A (MAX HARRY WEIL ET AL.), 28 Sept 1999 (28.09.99), column 10, line 1 - line 45, figures 7-10, abstract --	1-10,14
A	US 5694943 A (JAMES E. BREWER ET AL), 9 December 1997 (09.12.97), column 9, line 8 - column 10, line 24, abstract --	1-10,14
A	US 5724983 A (HARRY P. SELKER ET AL), 10 March 1998 (10.03.98), figure 1, abstract -- -----	1-10,14
<input type="checkbox"/> Further documents are listed in the continuation of Box C. <input checked="" type="checkbox"/> See patent family annex.		
* Special categories of cited documents: <div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <p>"A" document defining the general state of the art which is not considered to be of particular relevance</p> <p>"E" earlier application or patent but published on or after the international filing date</p> <p>"I" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)</p> <p>"O" document referring to an oral disclosure, use, exhibition or other means</p> <p>"P" document published prior to the international filing date but later than the priority date claimed</p> </div> <div style="width: 48%;"> <p>"I" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention</p> <p>"X" document of particular relevance: the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone</p> <p>"Y" document of particular relevance: the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art</p> <p>"&" document member of the same patent family</p> </div> </div>		
Date of the actual completion of the international search		Date of mailing of the international search report
18 January 2001		23 -01- 2001
Name and mailing address of the ISA: Swedish Patent Office Box 5055, S-102 42 STOCKHOLM Facsimile No. +46 8 666 02 86		Authorized officer Cilla Lyckman/mj Telephone No. +46 8 782 25 00

INTERNATIONAL SEARCH REPORT

International application No.
PCT/NO00/00289

Box I Observations where certain claims were found unsearchable (Continuation of item 1 of first sheet)

This international search report has not been established in respect of certain claims under Article 17(2)(a) for the following reasons:

1. ☐ Claims Nos.:
because they relate to subject matter not required to be searched by this Authority, namely:

2. ☐ Claims Nos.:
because they relate to parts of the international application that do not comply with the prescribed requirements to such an extent that no meaningful international search can be carried out, specifically:

3. ☐ Claims Nos.:
because they are dependent claims and are not drafted in accordance with the second and third sentences of Rule 6.4(a).

Box II Observations where unity of invention is lacking (Continuation of item 2 of first sheet)

This International Searching Authority found multiple inventions in this international application, as follows:

see extra sheet

1. ☐ As all required additional search fees were timely paid by the applicant, this international search report covers all searchable claims.
2. ☐ As all searchable claims could be searched without effort justifying an additional fee, this Authority did not invite payment of any additional fee.
3. ☐ As only some of the required additional search fees were timely paid by the applicant, this international search report covers only those claims for which fees were paid, specifically claims Nos.:

4. ☒ No required additional search fees were timely paid by the applicant. Consequently, this international search report is restricted to the invention first mentioned in the claims; it is covered by claims Nos.:

1-10, 14

Remark on Protest

- ☐ The additional search fees were accompanied by the applicant's protest.
☐ No protest accompanied the payment of additional search fees.

INTERNATIONAL SEARCH REPORT

International application No.
PCT/NO00/00289

The application is not considered to comply with the requirements of unity of invention for the following reasons:

According to PCT Article 34(3)(a-c) and Rule 13.2, an international application shall relate to one invention only or to a group of inventions linked by one or more of the same or corresponding "special technical features". "Special technical features" are features that define a contribution, which each of the claimed inventions, considered as a whole, makes over the prior art. If the novelty search reveals that the common technical features are known or do not involve an inventive step, a posterior lack of unity arises.

The invention claimed in claim 1 relates to a system for calculation of the probability of a successful defibrillatory shock. The system measures bio-electrical signals from electrodes connected to a patient.

The document US 5957856 A is considered to be of particular relevance; the claimed invention according to claim 1 cannot be considered novel or cannot be considered to involve an inventive step.

Hence, there remain 6 inventions as follows:

1. The system according to claim 2 including the additional feature that the bio-electric signals are ECG-signals.
2. The system according to claims 3-7 including the additional feature that the analysing unit is adapted to calculate the probability of success by means of an algorithm.
3. The system according to claims 8 and 9 including the additional feature that the output from the analysing unit is connected to a receiver.
4. The system according to claims 10 and 14 including the additional feature that the calculated probability of success is used in an algorithm for advisory support when a therapy form is selected.
5. The system according to claims 11-13 including the additional feature that the analysing unit identifies time periods with positive changes in the probability of success and parameters concerning the therapy and supplies the data to a receiver.
6. The system according to claim 15 with the additional feature that an arrangement for patient-specific data is connected to the analysing unit.

Since no common or corresponding special feature exists, no technical relationship can be identified between the different inventions.

INTERNATIONAL SEARCH REPORT
Information on patent family members

International application No.
PCT/NO 00/00289

Patent document cited in search report			Publication date	Patent family member(s)	Publication date
US	5957856	A	28/09/99	NONE	
US	5694943	A	09/12/97	US 5555888	17/09/96
US	5724983	A	10/03/98	US 5501229	26/03/96
				US 5718233	17/02/98